



GEO SAVER D

It allows the user to view the ECG and everything needed to know about the patient and ongoing rescue treatment on a 5.7" full-colour display compliant to IEC/EN 60601-2-27.

ECG can also be sent in real time to the AMISAVERCLOUD PLATFORM to enable another operator to monitor the ECG simultaneously with the local rescuer.



Defibrillator

Operation:	AED Semi-Automatic ; ECG Monitoring capability
Energies:	Standard max 200J or Power max 360J
Waveform:	Adaptive BTE (biphasic truncated exponential) conforming to patient chest's impedance
Protocols:	Various adult shock protocols available on request
Factory default:	Adult Standard escalating 150, 200, 200J Adult Power escalating 200, 250, 360J Pediatric (Standard or Power) 50J fixed
Charging time:	≤9 seconds with a new and fully charged battery depleted battery will result in a longer charging time
Analysis time:	IEC/EN 60601-2-4 from 4 to 15 seconds
Impedance:	20-200 ohms
Sensitivity:	IEC/EN 60601-2-4 (AHADB, MITDB source), 97%
Specificity:	IEC/EN 60601-2-4 (AHADB, MITDB source), 99%
Controls:	2 buttons: ON/OFF, shock button, and 3 buttons to surf the menu.
Indicators:	Status LED indicator informing on device condition Battery gauge with remaining capacity rate Audible alerts and text display with service alarms
Upgradeable:	through a USB cable, memory card or by remote through AMISAVERCLOUD

ECG Monitoring

Operations:	Through defibrillation pads or standard ECG electrodes attached to a separate 2-Lead patient monitoring reusable cable SAV-C0017
ECG size:	Manual setting through the menu
Heart Rate:	30-300 bpm
Sweep Speed:	25 mm/sec
Standard:	IEC/EN 60601-2-27 less then the points 202.6.2.101; 201.12.1.101.12,13; 208.6.6.2.101 not performed for the intended use of the device, as it is not intended for environments such as operating theatres or intensive care units
Display:	5,7" TFT colour, 640 x 480 pixel

Battery options

Type:	Li-SOCI2 Disposable, code SAV-C1032
Autonomy:	250 complete rescue cycles (shocks at 200J and CPR) or 160 complete rescue cycles (shocks at 360J and CPR) or 24 hours ECG monitoring for a new and fully charged battery(*) when stored in original packaging 5 years (*)
Shelf-Life:	4 years once installed to AED, assuming one battery insertion test and daily self-test but without switching AED on (*)
Battery-Life:	Li-ion Accumulator, code SAV-C1033
Recharging time:	2,5 hours with the charger station code SAV-C1034 (*) (recommended to charge every 4 months at least)
Autonomy:	200 shocks at 200J or 110 shocks at 360J or 14 hours in ECG Monitoring for a new fully charged accumulator (*)
Battery-Life:	2 years or 300 charging cycles (*)

Pads options

Type:	Disposable, pre-gelled and self-adhesive
Adult:	Code SAV-C0846, for patient >8 years or >25 kg
Pediatric:	Code SAV-C0016, for patient <8 years or <25 kg
Cable length:	120 cm
Shelf-Life:	30 months

Event recording

Internal memory:	up to 6 continuous hours of ECG and rescue events
Optional memory:	Removable SD card; length of storage depends on card capacity: a 2GB card records up to 100 hours "AED1LOG" text file with detailed self-test activity "AEDFILES" multimedia files with complete recorded information "Saver View Express" data manager software
Data recording:	
Event review:	

Physical

Size:	29,5 x 23 x 11,5 cm
Weight:	+/- 2,85 kg

Environmental

Operating temperature:	0°C to 55°C (32°F TO 131°F)
Storing/Shipping temperature:	-40°C to 70°C (-40°F TO 158°F) without battery
Humidity:	10% to 95% relative humidity non condensing
Sealing (IP Protection):	IEC/EN 60529 class IP56; splash proof, dust protected
Shock/Drop Abuse Endurance:	IEC/EN 60601-1 clause 21; 1 meter drop, impact, force, rough handling, mobile tolerance
Electrostatic Discharge:	IEC/EN 61000-4-2
Electromagnetic Compatibility:	IEC/EN 60601-1-2 Emission, Immunity
Electrical Protection:	IEC/EN 60601-1; Internally Powered, Type BF/CF
Directive 93/42/CEE and 2007/47/CE:	Class IIb
Radio Equipment Directive (RED):	Directive 2014/53/UE

Geoloc module

Frequency:	GSM: 850, 900, 1800, 1900 MHz; UMTS: 900, 2100 MHz GPS: 1575, 1600 MHz
------------	---

Geoloc battery options:

Type:	Li-SOCI2 Disposable, code SAV-C1038
Shelf-Life:	when stored in original packaging 5 years(*)
Battery-Life:	4 years once installed to AED, assuming one battery insertion test and daily self-test but without switching AED on(*)
Type:	Li-ion Accumulator, code SAV-C1039
Recharging time:	2,5 hours with the charger station code SAV-C1040 (*) (recommended to charge every 4 months at least)
Battery-Life:	2 years or 300 charging cycles (*)

(*)Temperature at 20°C Humidity 45% non-condensing